



# CERTIFICATE OF ACCREDITATION



## HI-Q Environmental Products Company, Inc.

7386 Trade Street  
San Diego, California 92121 USA

### Certificate of Accreditation Supplement ISO/IEC 17025:2017

Date of Issue: June 26, 2025 – Date of Expiration: June 25, 2028  
Certificate Number: AGS-US063022-1/3

Model	Range	Calibration & Measurement Uncertainty (%)	Reference Standard or Equipment	Methods
High Flow Air Sampler	5-50 CFM 140-1400 LPM	1.4 - 2.1%	High Flow Laminar Flow Element	PR-05
High Flow Air Flow Calibrator	5-50 CFM 140-1400 LPM	1.4 - 2.1%	High Flow Laminar Flow Element	PR-05
Medium Flow Air Sampler	0.5 - 30 CFM 15 - 850 LPM	1.4 %	Medium Flow Laminar Flow Element	PR-05
Medium Flow Air Flow Calibrator	0.5 - 30 CFM 15 - 850 LPM	1.4 %	Medium Flow Laminar Flow Element	PR-05
Low Flow Air Sampler	0.3 - 3 CFM 8 - 85 LPM	1 %	Low Flow Laminar Flow Element	PR-05
Low Flow Air Flow Calibrator	0.3 - 3 CFM 8 - 85 LPM	1 %	Low Flow Laminar Flow Element	PR-05
Very Low Flow Air Sampler	0.3 - 3 LPM	0.8 %	Very Low Flow Laminar Flow Element	PR-05
Very Low Flow Air Flow Calibrator	0.3 - 3 LPM	0.8 %	Very Low Flow Laminar Flow Element	PR-05